CERTIFICATE OF CALIBRATION

Model : 1/4" **Serial No :** 004643

Wrench Capacity : 25 **Inspector :**

Date of Calib. : 2013/01/07 **Unit** N.m

Set Point	*(CW) ±4%*		Actual Readings		
	Min	**Max**			
5	**4.8**	**5.2**	5.18	5.15	5.17
15	**14.4**	**15.6**	15.43	14.57	14.59
25	**24**	**26**	25.55	25.43	25.36

Testing Standard No : DIN-ISO-6789

Testing is in compliance with international standards with test equipment and reference standards calibrated by a laboratory traceable to international standards.

Tester Model : T-50

Serial No : 90355

The uncertainty of the equipment at 20℃ is < ±1%

Inspector Approval:

Manager
H.H.W.